

Щелочная фосфатаза, ALP

DGKC



Параметры для ввода в программу анализатора Architect c8000 / c4000

Configure assay parameters...

☒ General ☐ Calibration ☐ SmartWash ☐ Results ☐ Interpretation

Assay: Type: Version:
Number: Assay availability: Date:
Run controls for onboard reagents by: Time:
Operator:

☒ Reaction definition ☐ Reagent / Sample ☐ Validity checks

Reaction mode:

Primary Secondary Read times
Wavelength: / Main: --
Last required read: Flex: --
Absorbance range: -- Color correction: --
Sample blank type:

Assay of

Done

Cancel

?

Configure assay parameters...

☒ General ☐ Calibration ☐ SmartWash ☐ Results ☐ Interpretation

Assay: Type: Version:
Number: Assay availability: Date:
Run controls for onboard reagents by: Time:
Operator:

☐ Reaction definition ☒ Reagent / Sample ☐ Validity checks

Reagent:
Diluent:
Diluent dispense mode:

R1 R2
Reagent volume:
Water volume:
Dispense mode:

Dilution name	Sample	Diluted sample	Diluent	Water	Dilution factor	Default dilution
Standart	: 3.2				= 1:1.00	<input checked="" type="radio"/>
1:20	: 5.0	3.2	95.0		= 1:20.00	<input type="radio"/>
1:10	: 10	3.2	90		= 1:10.00	<input type="radio"/>

Assay of

Done

Cancel

?

Щелочная фосфатаза, ALP

DGKC



Параметры для ввода в программу анализатора Architect c8000 / c4000

Configure assay parameters...

General Calibration SmartWash Results Interpretation

Assay: ALP Assay number: *** Date: 20.02.2020
Calibration method: Linear Time: 17:14
Operator: FSE

Calibrators Volumes Intervals Validity checks

Calibrator set: TruCal_U
Replicates: 2 [Range 1 - 3]

Calibrator level: Blank: Water
Cal 1: TruCal_U
Cal 2: None
Cal 3: None
Cal 4: None
Cal 5: None
Cal 6: None

Concentration: 0.0000

Assay of

Done Cancel

?

Configure assay parameters...

General Calibration SmartWash Results Interpretation

Assay: ALP Assay number: *** Date: 20.02.2020
Calibration method: Linear Time: 17:14
Operator: FSE

Calibrators Volumes Intervals Validity checks

Calibrator:

Calibrator level	Sample	Diluted sample	Diluent	Water
Blank: Water	3.2			
Cal 1: TruCal_U	3.2			
Cal 2: None				
Cal 3: None				
Cal 4: None				
Cal 5: None				
Cal 6: None				

Assay of

Done Cancel

?

Щелочная фосфатаза, ALP

DGKC



Параметры для ввода в программу анализатора Architect c8000 / c4000

Единицы измерения - Ед/л; знаков после запятой 1 (**Configure result units**).

Диапазон метода: 2-4500 Ед/л (Configure assay parameters\ Results).

Референсный интервал: 0-258 Ед/л (Configure assay parameters\ Results).

При наличии ЛИС, не обязательно.

Диапазон нормальных значений указан в соответствии с рекомендациями производителя реагентов.

Калибровка линейная, калибратор TruCal U, нулевой - вода (или физраствор). Или по фактору Контроль по TruLab N и TruLab P.